patient commission/configuration ID _____



Item	20mm	Description	Potential Problem	Measure	Recommended Inspection, Potential Replacement*	Latest Replacement after
System Joint without Feather Keys						
1	SB1069-L0960	bearing nut	wear	replacing	every 6 months	36 months
5	GS2611-040	sliding washer, Ø = 26mm	wear	replacing	every 6 months	18 months
5	GS2611-045					
5	GS2611-050					
5	GS2611-055					
5	GS2611-060					
6	SF0765-3L/AL	cover plate, left lateral or right medial**	wear	replacing	every 6 months	36 months
6	SF0765-3R/AL	cover plate, left medial or right lateral**				
8	FE2966-01	coil spring, golden**	wear	replacing	every 6 months	18 months
9	SF0845-10	pin**	wear	replacing	every 6 months	36 months
11	SC1416-L14	countersunk flat head screw with hexalobular socket**	wear	replacing	every 6 months	36 months
12	SC1416-L14	countersunk flat head screw with hexalobular socket (axle screw)**				
6-12	SF7985-L/AL	functional unit plug + go modularity, left lateral or right medial***	loss of function, wear	replacing	every 6 months	36 months
6-12	SF7985-R/AL	functional unit plug + go modularity, left medial or right lateral***				
System Stirrups, System Anchors and System Side Bars						
13	BR1211-L030	sliding bushing	wear	replacing	every 6 months	18 months
14	see catalogue	system stirrup	wear or breakage	replacing	every 6 months	48 months
15	see catalogue	system anchor and system side bar	wear or breakage	replacing	every 6 months	if required

* depending on the assessment of the distributor of the custom-made product regarding the patient's usage behaviour
** is part of the functional unit
*** included system components can be exchanged separately

WP0057-GB-2024-08