

Instructions for Use for Qualified Specialists in Orthopaedic Technology System Ankle Joints

EN



NEURO SWING Carbon



NEURO CLASSIC Carbon

Content	Page
1. Information	4
2. Safety Instructions	4
2.1 Classification of the Safety Instructions	4
2.2 All Instructions for a Safe Handling of the System Ankle Joint	5
3. Use	7
3.1 Intended Use	7
3.2 Indication	7
3.3 Contraindication	7
3.4 Qualification	7
3.5 Application	7
3.6 Product Range	7
3.7 Combination Possibilities with Other System Joints	8
4. Joint Function	8
5. Scope of Delivery	8
6. Load	9
7. Tools for Assembling the System Joint	9
8. Mounting the System Joint	9
8.1 Mounting the System Stirrup	9
8.2 Checking the Free Movement	10
8.3 Mounting the Spring Units NEURO SWING Carbon	10
8.4 Securing the Screws	10
9. Adjustment Options on the Orthosis	11
9.1 Adjustments on the Spring Unit NEURO SWING Carbon	11
9.1.1 Adjustable Alignment NEURO SWING Carbon	11
9.1.2 Variable Spring Force NEURO SWING Carbon	12
9.2 Reading the Joint Angles	12
10. Connecting to the System Side Bar/System Anchor	12
11. Conversion of the Orthosis with NEURO SWING Carbon System Ankle Joint	13
12. Maintenance	13
12.1 Documentation of Maintenance in the Orthosis Service Passport	14
12.2 Replacing the Sliding Washers	14
12.3 Dirt Removal	14
13. Period of Use	14
14. Storage	15

.....

15. Spare Parts	15
15.1 Exploded View Drawing NEURO SWING Carbon	15
15.2 Spare Parts for All System Ankle Joints	16
15.3 Spare Parts for the NEURO SWING Carbon System Ankle Joint	16
15.4 Spring Units and Static Dorsiflexion Stop NEURO SWING Carbon	17
15.5 Spare Parts for the NEURO CLASSIC Carbon System Ankle Joint	17
16. Disposal	17
17. Signs and Symbols	18
18. CE Conformity	18
19. Legal Information	18
20. Information for the Treatment Documentation	19
21. Handing Over the Orthosis	20

1. Information

These instructions for use are addressed to qualified specialists in orthopaedic technology and do not contain any notes about dangers which are obvious to them. To achieve maximum safety, please instruct the patient and/or care team in the use and maintenance of the product.






For a simplified illustration, all basic work steps are shown with the **NEURO SWING Carbon** system ankle joint (fig. 1) as an example. They can be transferred to all mentioned system joints.



fig. 1

2. Safety Instructions

2.1 Classification of the Safety Instructions

 DANGER	Important information about a possible dangerous situation which, if not avoided, leads to death or irreversible injuries.
 WARNING	Important information about a possible dangerous situation which, if not avoided, leads to reversible injuries that need medical treatment.
 CAUTION	Important information about a possible dangerous situation which, if not avoided, leads to light injuries that do not need medical treatment.
NOTICE	Important information about a possible situation which, if not avoided, leads to damage of the product.

All serious incidents according to Regulation (EU) 2017/745 which are related to the product have to be reported to the manufacturer and to the competent authority of the Member State in which the qualified specialist in orthopaedic technology and/or the patient is established.

2.2 All Instructions for a Safe Handling of the System Ankle Joint

DANGER

Potential Traffic Accident Due to Limited Driving Ability

Advise the patient to gather information about all safety and security issues before driving a motor vehicle with orthosis. The patient should be able to drive a motor vehicle safely.

WARNING

Risk of Falling Due to Improper Handling

Inform the patient about the correct use of the system joint and possible dangers, especially with regards to excessive mechanical stress (e.g. due to sports, increased activity or weight gain). Also inform the patient that the system joint may only be demounted and maintained by a qualified specialist in orthopaedic technology. Any handling of the system joint and the orthosis by the patient that goes beyond the tasks described in the instructions for use for patients is not permitted.

WARNING

Risk of Falling Due to Improper Processing

Process the system joint according to the information in these instructions for use. Deviating processing and modifications of the system joint require the written consent of the manufacturer.

WARNING

Risk of Falling Due to Loosening of the Bearing Nut

Secure the screw of the joint case with the specified torque and the corresponding adhesive and make sure that no sliding washers are damaged in the process.

WARNING

Risk of Falling Due to Incorrectly Selected System Components

Make sure that the system joint and the system components are not overloaded and are functionally adapted to the requirements and needs of the patient in order to avoid joint dysfunction.

WARNING

Risk of Falling Due to Permanent Higher Load

If patient data has changed (e.g. due to weight gain, growth or increased activity), recalculate the expected load on the system joint, plan the treatment again and, if necessary, produce a new orthosis.

WARNING

Risk of Falling Due to Improper Shoe/Wrong Shoe Pitch

Advise the patient to wear a shoe to which the orthosis is adjusted in order to avoid joint dysfunction.

WARNING

Risk of Falling Due to Excessive Readjustment of the Spring Unit

Adjust the spring unit according to the information in these instructions for use. Do not make readjustments of more than 10°. Use the laser markings on the system stirrup and the joint case to check the readjustment.

WARNING

Damage to the Anatomical Joint Due to Incorrect Position of the Joint's Mechanical Pivot Point

Determine the joint's mechanical pivot points correctly in order to avoid a permanent incorrect load on the anatomical joint. Please refer to the online tutorials on the FIOR & GENTZ website or contact Technical Support.

WARNING

Jeopardising the Therapy Goal by Not Providing the Necessary Free Movement

Check if the system joint moves freely in order to avoid restrictions of the joint function. Use suitable sliding washers according to the information in these instructions for use.

WARNING

Jeopardising the Therapy Goal Due to Incorrectly Adjusted Spring Units

Screw in the spring unit up to the system stirrup and do not preload the spring unit. If the stops are reached too early or too late, either the range of motion is restricted or the patient is not sufficiently stabilised by the orthosis, which worsens the gait. In order to utilise the full functional potential of the orthosis, the spring units must be suitably selected and correctly adjusted.

NOTICE

Limitation of the Joint Function Due to Improper Processing

Errors in processing can impair the joint function. Pay particular attention to:

- correctly connecting the system side bar/system anchor with the joint case in accordance with the production technique;
- greasing the joint components only slightly and
- adhering to the maintenance intervals.

NOTICE

Limitation of the Joint Function Due to Improper Dirt Removal

Inform the patient on how to properly remove dirt from the orthosis and the system joint.

NOTICE

Limitation of the Joint Function Due to Lack of Maintenance

Respect the specified maintenance intervals in order to avoid joint dysfunction. Also inform the patient about the maintenance appointments to be respected. Enter the next maintenance appointment in the orthosis service passport of the patient.

3. Use

3.1 Intended Use

The FIOR & GENTZ system ankle joints must be used exclusively for the orthotic treatment of the lower extremity. The system joints are only allowed to be used for producing an AFO or a KAFO. Every system joint influences the orthosis' function and thus also the function of the leg. The system joint may only be used for one fitting and must not be reused.

3.2 Indication

The indications for the treatment with an orthosis for the lower extremity are insecurities that lead to a pathological gait. This can be caused, for example, by paralyses, structurally conditioned deformities/ malfunctions or as a result of physical trauma and/or surgery.

The physical conditions of the patient, such as muscle strength or activity level, are crucial for the orthotic treatment. An evaluation regarding the safe handling of the orthosis by the patient must be carried out.

All system ankle joints can also be used for the prosthetic treatment of patients with partial foot amputations. For this purpose, the orthosis produced for the patient by a qualified specialist in orthopaedic technology (custom-made product) is combined with a foot prosthesis. Further information can be found in the **Guide to Partial Foot Amputations** (see QR code, fig. 2).



fig. 2

3.3 Contraindication

The system joint is not suitable for treatments that were not described in paragraph 3.2, such as a treatment of the upper extremity or a treatment with a prosthesis or ortho-prosthesis that affects more than just part of the foot, for example after amputations of leg segments.

3.4 Qualification

The system joint must only be handled by a qualified specialist in orthopaedic technology.

3.5 Application

All FIOR & GENTZ system joints were developed for everyday life activities such as standing and walking. Extreme impact stress, which occurs for example during long jump, climbing and parachuting, is excluded. The carbon system ankle joints are water-resistant and therefore suitable for use in wet areas. They are equipped with a water-resistant carbon fibre reinforced joint case and a seawater-resistant stainless steel screwing. The **NEURO SWING Carbon** also has spring units which are inside water- and dirt-resistant spring unit sleeves. The spring units of the system joint are waterproof in depths of up to 3 metres. The system joints can be used at a maximum temperature of +60°C.

3.6 Product Range

These instructions for use provide information on the following system ankle joints:



NEURO SWING Carbon



NEURO CLASSIC Carbon

3.7 Combination Possibilities with Other System Joints

The carbon system ankle joints can be mounted in combination with carbon system knee joints from the FIOR & GENTZ product range in a water-resistant orthosis. The combination with other system knee joints from the FIOR & GENTZ product range is also possible. The **NEURO CLASSIC Carbon** can be used as a supporting joint for the **NEURO SWING Carbon**.

We recommend that you use the Orthosis Configurator when selecting all system components for your orthosis and follow the recommendations of the configuration result.

4. Joint Function

Due to the spring units used, the **NEURO SWING Carbon** system ankle joint has the following functions:

System Component	Function
spring units	dorsal (posterior spring unit): - determination of the maximum range of motion in plantar flexion - integrated dorsiflexion assist - controlled lowering of the foot during loading response
	ventral (anterior spring unit): - determination of the maximum range of motion in dorsiflexion - increased energy return during heel lift to support push off
	ventral (static dorsiflexion stop): - limitation of the maximum range of motion in dorsiflexion
	dorsal and ventral: - dynamically bringing the patient from a bent into an upright position as well as improving the patient's stability while walking and standing by balancing the body

5. Scope of Delivery

Description	Quantity
system ankle joint (without figure)	1
set 2-component adhesive with primer (fig. 3)	1
orthosis joint grease, 3g (without figure)	1
assembly/lamination dummy (fig. 4)	1

The corresponding spring units and system stirrups have to be ordered separately.



fig. 3



fig. 4

6. Load

The actual load on the system joints is based on the relevant patient data. The load and the appropriate system components can be determined by using the Orthosis Configurator. We recommend that you use the system components determined by the Orthosis Configurator when producing an orthosis and mind the recommended production technique. You will find information on the production techniques in the section "Online Tutorials" on the FIOR & GENTZ website.

7. Tools for Assembling the System Joint

Tools	System Width			
	12mm	14mm	16mm	20mm
T15 hexalobular screwdriver/bit	x	-	-	-
T20 hexalobular screwdriver/bit	-	x	x	x
torque screwdriver, 1–6Nm	x	x	x	x
hexagonal screwdriver with spherical head, 4 x 100mm	x	-	-	-
hexagonal screwdriver with spherical head, 5 x 100mm	-	x	x	x
sliding washer centring pin	x	x	x	x

8. Mounting the System Joint

The system joint is delivered fully assembled. All functions are checked beforehand. You have to disassemble the system joint for mounting it to the orthosis and for maintenance. To ensure an optimal functioning, follow the assembly instructions below. Secure the screw with the torque specified in paragraph 8.4.

You can find more information on the assembly in the online tutorial **Joint Assembly NEURO CLASSIC Carbon, NEURO SWING Carbon** (see QR code, fig. 5) on the FIOR & GENTZ website.

In the following, the assembly is illustrated with the **NEURO SWING Carbon** system ankle joint as an example.



Only use the FIOR & GENTZ orthosis joint grease to grease the system components.



fig. 5

8.1 Mounting the System Stirrup

- 1 Before the assembly, clean the thread of the bearing nut with LOCTITE® 7063 Super Clean. Allow the thread to air-dry for 10 minutes.
- 2 Grease the sliding surfaces of the bearing nut and, if existing, the contact surfaces of the system stirrup between system stirrup and spring units with orthosis joint grease.
- 3 Grease the two sliding washers **slightly** on both sides with orthosis joint grease.

- 4 Place the sliding washers onto both sides of the system stirrup (fig. 6).
- 5 Slide the system stirrup from below into the joint case (fig. 7). Make sure that the sliding washers remain in the correct position. To do so, use the sliding washer centring pin.



fig. 6



fig. 7




fig. 8



fig. 9



fig. 10



Make sure not to damage the sliding washers during the assembly. Jammed sliding washer particles can cause lateral play in the system joint.

- 6 Put the bearing nut into the joint case. The bearing nut must be fully inserted in the opening (fig. 8).
- 7 Place the cover disc onto the joint case's front.
- 8 Screw in the countersunk flat head screw (S1; fig. 9).

8.2 Checking the Free Movement

Tighten the screw for the joint case with the appropriate torque (see paragraph 8.4). Check if the system joint moves freely. If the system joint runs with lateral play, mount the next thicker sliding washer. If it does not move freely (it is jammed), mount the next thinner sliding washer.

8.3 Mounting the Spring Units NEURO SWING Carbon

For the **NEURO CLASSIC Carbon** system ankle joint, skip these steps and continue the assembly at paragraph 8.4.

- 1 Screw the spring unit for dorsiflexion or the static dorsiflexion stop into the anterior spring duct until the required alignment of the orthosis is achieved (fig. 10).
- 2 Screw the spring unit for plantar flexion into the posterior spring duct until it touches the system stirrup. Do not preload the spring unit.



Do not disassemble the spring unit as it is under pressure. There is a risk of injury when opening the spring unit sleeve. The spring unit and the O-ring for the **NEURO SWING Carbon** system ankle joint must not be greased.

8.4 Securing the Screws

The screws are secured after the orthosis has been produced and tried on and before it is handed over to the patient.

- 1 Loosen the screw for the joint case (fig. 9) after checking the free movement and remove it from the joint case.
- 2 Apply a small drop of LOCTITE® 243 medium strength to the thread of the screw.

- 3 Secure the screw for the joint case (fig. 9) with the torque corresponding to the system width.
- 4 Let the adhesive harden (final strength after approx. 24 hours).

Screw for Joint Case	System Width			
	12mm	14mm	16mm	20mm
S1 (screw 1, axle screw)	3Nm	4Nm	4Nm	4Nm



The screw of the joint case is not secured with the necessary torque at delivery. You can also find information on the torque on the cover disc of the system joint.

9. Adjustment Options on the Orthosis

The orthosis can be individually adapted to the patient's needs with adjustable system ankle joints (fig. 11). The adjustments described do not influence each other and can be made independently of each other.




fig. 11



Mind the correct adjustment of the dorsiflexion stop when mounting the system ankle joint. It is decisive for the entire alignment of the orthosis. You can find more information on this in the online tutorial **AFO Alignment Guidelines** (see QR code, fig. 12) on the FIOR & GENTZ website.



fig. 12

9.1 Adjustments on the Spring Unit NEURO SWING Carbon

There are spring units with disc springs (2) and coil springs (3). The alignment of the orthosis can be adjusted by screwing and unscrewing the spring units (1) (fig. 11). The spring force can be changed with spring units in different strengths.

9.1.1 Adjustable Alignment NEURO SWING Carbon

Always unscrew only one spring unit at a time to adjust the angle between lower leg and foot (fig. 13). Only then, screw in the other spring unit until it touches the system stirrup. Do not preload the spring unit as this will restrict the maximum possible range of motion. An O-ring is attached to the external thread of the spring unit to ensure that the position of the spring unit does not change.




fig. 13

9.1.2 Variable Spring Force NEURO SWING Carbon

The spring force can be changed by exchanging the spring units. Insert a spring unit into the spring duct that corresponds with the required spring force. There are five spring units with spring forces ranging from normal to extra strong (fig. 14). The static dorsiflexion stop (fig. 14) does not have any dynamic properties (0° range of motion). Note that the spring unit as well as the static dorsiflexion stop determine the maximum possible range of motion.




fig. 14

9.2 Reading the Joint Angles

There are markings (fig. 15) on the joint case and the system stirrup which indicate the angle of the system components to each other. This allows you to check the individual normal posture (the orthosis' basic alignment), record the joint angle and compare later deviations. The joint angle in the individual normal posture must not be outside the degree markings.

The distances between the degree markings for each system width can be seen in the following table.

Degree Marking				
System Width	12mm	14mm	16mm	20mm
Degree	5°	2°	2°	2°




fig. 15

10. Connecting to the System Side Bar/ System Anchor

The system side bar/system anchor must be adhered to the system joint according to the recommended production technique (fig. 16). It is adhered after the orthosis parts have been tempered. Before using the adhesive set, make sure that the expiry date has not yet passed. The adhesive set should be stored in a cool place.



fig. 16



Note that the orthosis should not be tempered after connecting the system side bar/system anchor to the system joint. The properties of the adhesive connection change at temperatures that are too high.

You can find more information in the Instructions for Use for Qualified Specialists in Orthopaedic Technology System Side Bars and System Anchors for Carbon System Joints (see QR code, fig. 17). You will find information on the production techniques in the section "Online Tutorials" on the FIOR & GENTZ website.



fig. 17

11. Conversion of the Orthosis with NEURO SWING Carbon System Ankle Joint

An orthosis with a NEURO SWING Carbon system ankle joint can be converted into an orthosis with a NEURO CLASSIC Carbon system ankle joint by exchanging the system joint.

12. Maintenance

Check the system joint regularly for wear and functionality. In particular, check the joint components listed in the following table for the possible problems described and, if necessary, take the appropriate measures. Also check the functionality after every maintenance carried out. It must be possible to move the system joint without problems or unusual noises. Make sure that there is no lateral play and no play around the axis.

Joint Component	Potential Problem	Measure	Recommended Inspection, Potential Replacement*	Latest Replacement
O-ring for securing the spring unit	wear	replacing O-ring	every 6 months	every 18 months
spring unit	wear	replacing spring unit	every 6 months	every 18 months
	noise of spring unit	replacing spring unit	every 6 months	every 18 months
sliding washer	wear	replacing sliding washer, see paragraph 12.2	every 6 months	every 18 months
sliding bushing	wear	replacing sliding bushing	every 6 months	every 18 months
countersunk flat head screw with hexalobular socket	wear	replacing countersunk flat head screw	every 6 months	every 36 months
bearing nut	wear	replacing bearing nut	every 6 months	every 36 months
system stirrup	wear or breakage	replacing system stirrup	every 6 months	every 48 months
adhesive bond (system side bar/ system anchor and system joint)	breakage of adhesive bond	adhering new carbon joint component	every 6 months	if required

* depending on the assessment of the distributor of the custom-made product regarding the patient's usage behaviour

Clean the thread of the bearing nut with LOCTITE® 7063 Super Clean at every maintenance. Allow the thread to air-dry for 10 minutes.

Secure the screw for the joint case with the appropriate torque and LOCTITE® 243 medium strength at every maintenance (see paragraph 8.4). Remove all adhesive residues first.



When disassembling the system joint, make sure to fix the bearing nut on the backside with one finger while unscrewing the screw. This prevents the bearing nut from slipping out of the opening and damaging the material of the joint case.



You can find the individual maintenance plans for system joints in the download area (see QR code, fig. 18) on the FIOR & GENTZ website.

fig. 18

12.1 Documentation of Maintenance in the Orthosis Service Passport

The patient receives an orthosis service passport (fig. 19) from a qualified specialist in orthopaedic technology when the orthosis is handed over. The orthosis must be checked regularly according to the specifications in the maintenance plan in order to maintain its function and to ensure the safety of the patient. The maintenance appointments are noted and confirmed in the orthosis service passport.



fig. 19

12.2 Replacing the Sliding Washers

Sliding washers are available in different thicknesses (e.g. GS1911-040 is 0.40mm thick). Each thickness has a different marking (fig. 20). You will find the article numbers of the premounted sliding washers on the back page of these instructions for use. Please use the sliding washer centring pin to position the sliding washers.




fig. 20

12.3 Dirt Removal

The system ankle joint is suited for use in wet areas. Nevertheless, dirt must be removed when necessary and during regular maintenance. For this purpose, disassemble the system joint – but not the spring units (if existing) – and clean the soiled parts as well as, if existing, the spring unit sleeves with a dry cloth.

In order to optimise the lifespan, we recommend rinsing the orthosis with clear tap water, especially after using it in salt water, chlorine water and sand.

13. Period of Use

To guarantee a safe use and complete functionality as well as an unlimited period of use of the system joints, you must adhere to the following conditions:

- Adhere to the specified maintenance intervals without interruption and document each maintenance (see paragraph 12).
- Adhere to the determined maintenance conditions (see paragraph 12).
- Check the wear parts, as required, and exchange them at the defined intervals (see paragraph 12).
- Check the adjustment of the system joint during maintenance and correct it if necessary (see paragraph 12).
- Check the functionality of the system joint during maintenance (see paragraph 12).
- The maximum load determined during the planning of the custom-made product shall not be exceeded by changes in the patient data (e.g. due to weight gain, growth or increased activity). If the determined maximum load on the system joints is exceeded, the system joint must no longer be used. When planning the custom-made product, expected changes in patient data need to be taken into account.
- The period of use of the water-resistant system joints can be affected by use in salt water, chlorine water or sand. After use in salt water, chlorine water or sand, rinse the system joint with clear tap water. Instruct the patient accordingly.

- The period of use of the system joints ends with the period of use of the custom-made product (orthosis).
- The multiple use of the system joint in another custom-made product is not allowed (see paragraph 19).

14. Storage

It is recommended to store the system joint in its original packaging until the custom-made product is produced.

15. Spare Parts

15.1 Exploded View Drawing NEURO SWING Carbon

The exploded view drawing of the NEURO SWING Carbon system ankle joint also serves as an exemplary illustration for the NEURO CLASSIC Carbon system ankle joint.



fig. 21

All system stirrups of the system ankle joints are delivered with an integrated sliding bushing.

15.2 Spare Parts for All System Ankle Joints

Item	Article Number for System Width				Description
	12mm	14mm	16mm	20mm	
1	SF0591-C/1	SF0592-C/1	SF0593-C/1	SF0595-C/1	bearing nut
3	GS1409-*	GS1911-*	GS2413-*	GS2815-*	sliding washer*
4	BR1009-L020	BR1211-L025	BR1312-L030	BR1514-L030	sliding bushing
5	SF0591-C/2	SF0592-C/2	SF0593-C/2	SF0595-C/2	cover disc
6	SC1404-L10	SC1405-L11	SC1406-L14	SC1406-L14	countersunk flat head screw with hexalobular socket

* Sliding Washers

Article Number for System Width				
12mm	14mm	16mm	20mm	
Ø = 14mm	Ø = 19mm	Ø = 24mm	Ø = 28mm	
GS1409-040	GS1911-040	GS2413-040	GS2815-040	
GS1409-045	GS1911-045	GS2413-045	GS2815-045	
GS1409-050	GS1911-050	GS2413-050	GS2815-050	
GS1409-055	GS1911-055	GS2413-055	GS2815-055	
GS1409-060	GS1911-060	GS2413-060	GS2815-060	

15.3 Spare Parts for the NEURO SWING Carbon System Ankle Joint

Item	Article Number for System Width				Description
	12mm	14mm	16mm	20mm	
2	SF0501-C	SF0502-C	SF0503-C	SF0505-C	joint case

15.4 Spring Units and Static Dorsiflexion Stop NEURO SWING Carbon

Item	Article Number for System Width				Description
	12mm	14mm	16mm	20mm	
7	VE3771-085/13	VE3771-100/12	VE3771-12/12	VE3771-140/15	O-ring for securing the spring unit
8	SF5801-C/15/03	SF5802-C/15/05	SF5803-C/15/07	SF5805-C/15/18	spring unit, blue, normal, max. 15° range of motion
8	SF5801-C/15/06	SF5802-C/15/11	SF5803-C/15/15	SF5805-C/15/25	spring unit, green, medium, max. 15° range of motion
8	SF5801-C/10/12	SF5802-C/09/16	SF5803-C/10/21	SF5805-C/10/40	spring unit, white, strong, max. 10° range of motion
8	SF5801-C/10/19	SF5802-C/10/29	SF5803-C/10/31	SF5805-C/10/60	spring unit, yellow, very strong, max. 10° range of motion
8	SF5801-C/05/33	SF5802-C/05/53	SF5803-C/05/63	SF5805-C/05/99	spring unit, red, extra strong, max. 5° range of motion
8	-	-	SF5803-C/0	SF5805-C/0	static dorsiflexion stop, black, 0° range of motion

15.5 Spare Parts for the NEURO CLASSIC Carbon System Ankle Joint

The assignment of the items as shown in the exploded view drawing of the NEURO SWING Carbon system ankle joint serves as guidance. The spare parts of the NEURO CLASSIC Carbon system ankle joint are not identical to the picture.

Item	Article Number for System Width		Description
	16mm		
2	SF0103-C		joint case

16. Disposal

Dispose of the system joint and its individual parts properly. The product must not be disposed of with the residual waste (fig. 22). Please comply with the applicable national laws and local regulations for the proper recycling of recyclable materials.



fig. 22



For proper disposal, it is necessary to demount the system joint from the orthosis.

17. Signs and Symbols



CE labelling according to Regulation (EU) 2017/745 for medical devices



medical device



article number



manufacturer



batch code



serial number



follow the instructions for use



single patient – multiple uses



Unique Device Identifier – product identification number

18. CE Conformity

We declare that our medical devices as well as our accessories for medical devices are in conformity with the requirements of Regulation (EU) 2017/745. Therefore, the FIOR & GENTZ products bear the CE marking.

19. Legal Information

With the purchase of this product, our General Terms and Conditions of Business Transactions, Sales, Delivery and Payment will apply. The warranty expires, for example, if the product is mounted several times. Please note that the product is not supposed to be combined with other components or materials than with those recommended in the configuration result of the FIOR & GENTZ Orthosis Configurator. The combination of the product with products from other manufacturers is not permitted.

The information in these instructions for use is valid at the date of printing. The contained product information serves as guidelines. Subject to technical modifications.

All copy rights, particularly the distribution, copy and translation of these instructions for use or any part of them, must be authorised by FIOR & GENTZ Gesellschaft für Entwicklung und Vertrieb von orthopädietechnischen Systemen mbH. Reprints, copies and any other electronic reproduction, even partial, must be authorised in writing by FIOR & GENTZ Gesellschaft für Entwicklung und Vertrieb von orthopädietechnischen Systemen mbH.

.....

20. Information for the Treatment Documentation

Add these instructions for use to your treatment documentation!

Patient Data

Name	
Address	
Postcode, City	
Home Telephone	
Telephone at Work	
Insurance	
Insurance No.	
Attending Physician	
Diagnosis	

21. Handing Over the Orthosis

The qualified specialist in orthopaedic technology has also handed over the instructions for use for patients as well as the orthosis service passport to you as a patient, parent or care team. By means of these instructions for use, the functions and handling of the orthosis were explained to you in detail. You will find the next maintenance appointment in the orthosis service passport. Bring the orthosis service passport with you to every maintenance appointment.



Place, Date

Signature Patient

Leg Side

left

right

Mounted Sliding Washers

1. GS _____ - _____

2. GS _____ - _____



PB1400-DE/GB-2025-02



FIOR & GENTZ

Gesellschaft für Entwicklung und Vertrieb
von orthopädietechnischen Systemen mbH

Dorette-von-Stern-Straße 5
21337 Lüneburg (Germany)

+49 4131 24445-0
+49 921 95659554

info@fior-gentz.de
www.fior-gentz.com